

**Radio-communication service tester
PCT-430**

CHECK METHOD
ИТЦК468166.002СМ

The present check method is for radio-communication service tester PCT-430 (hereafter referred to as “tester”). It prescribes procedures and instruments of primary and periodical checks according to ИТЦК468166.002ТУ.

The tester is checked for determination of its operational suitability:

- primary check, after manufacturing or repair;
- periodical check, during operation;
- exceptional check, in the case of loss of check documents; after long-term storage of tester.

The tester should be checked in accordance with the prescribed procedure.

Check interval is 1 year.

Metrological performances of the tester are given in appendix A.

1 CHECK STAGES

1.1 For the check, the operations given in Table 1 should be carried out:

Table 1

Name of operation	Item number	Check	
		Primary check	Periodical check
1 External examination	6.1	Yes	Yes
2 Testing	6.2	Yes	Yes
3 Determination of metrological performances:			
3.1 output signal frequency forming relative error estimation (high-frequency)	6.3.1	Yes	Yes
3.2 output voltage setting absolute error estimation (high-frequency)	6.3.2	Yes	Yes
3.3 signal frequency deviation setting relative error estimation (high-frequency)	6.3.3	Yes	Yes
3.4 output signal frequency forming relative error estimation (low-frequency)	6.3.4	Yes	Yes
3.5 output voltage setting absolute error estimation (low-frequency)	6.3.5	Yes	Yes
3.6 signal frequency measuring relative error estimation (high-frequency)	6.3.6	Yes	Yes
3.7 signal frequency deviation measuring relative error estimation (high-frequency)	6.3.7	Yes	Yes
3.8 output signal power measuring relative error	6.3.8	Yes	Yes

estimation (high-frequency)

3.9 signal frequency measuring absolute error estimation (low-frequency) 6.3.9 Yes Yes

3.10 alternating voltage measuring relative error estimation 6.3.10 Yes Yes

3.11 alternating voltage distortion factor measuring relative error estimation 6.3.11 Yes Yes

3.12 direct voltage measuring relative error estimation 6.3.12 Yes Yes

1.2 If the result is negative during one of the operations, the check will be stopped.

2 CHECK INSTRUMENTS

2.1 For the check, the measuring instruments and auxiliary equipment indicated in the Table 2 must be used.

Table 2

Item number of check method	Name, type, model of standard instrument or auxiliary equipment	Key specifications and/ or metrological performances
1	2	3
6.3.1	Spectrograph C4-74	Frequency measurement in the range from 300 Hz to 300 MHz. $\Delta_{\text{abs}} = \pm(1 \cdot 10^{-7} \cdot f_c + 1 / t_{\text{cu}})$, Hz. Voltage level measurement in the range from 300 nV to 3 V. $\delta_{\text{rel}} = \pm 10 \%$.
6.3.2	Programmable high-frequency signal generator $\Gamma 4-164$	Frequency range from 0.1 to 639.999 MHz. $\delta_{\text{rel}} = \pm 0,5 \cdot 10^{-5} \%$. Output voltage range from $0,032 \cdot 10^{-6}$ to 2 V. $\Delta_{\text{abs}} = \pm 0,1 \text{ dB}$.
	Spectrograph C4-74 Selective microvoltmeter SMV8.5	
6.3.3	Low-frequency signal generator $\Gamma 3-112/1$	Output voltage: $0 \div 25 \text{ V}$. Frequency range: from 10 Hz to 1 MHz. $\Delta_{\text{rel}} = \pm 6 \%$. Peak value measurement limits:

		0.1 ÷ 1000 kHz.
	Computing modulation	$\Delta_{a\delta c} = \pm (0.1\Delta f + 12 \cdot 10^{-8} f_c + 0.005)$, kHz.
	meter CK3-45	Mean square value measurement limits: 0.005 ÷ 300 kHz. $\Delta_{a\delta c} = \pm (0.1\Delta f + 4 \cdot 10^{-8} f_c + 0.002)$, kHz.
	Selective microvoltmeter	
	SMV8.5	
6.3.4	cymometer Ч3-54	Frequency measurement limits: from 0.1 to 120 MHz. $\Delta_{rel} = \pm \{5 \cdot 10^{-7} + 1 / (f_{chan} \cdot t_{cq})\}$
6.3.5	Universal digital voltmeter B7-34	Sinusoidal voltage measurement range up to 500 V in frequency range from 20 Hz to 100 kHz $\Delta_{rel} = \pm \{0.15 + 0.05[(U_{кx}/U_x) - 1]\}$, % Distortion factor measurement range: 0.03 ÷ 100 %.
	Distortion meter C6-12	$\Delta_{abs} = \pm (0.05 K_r + 0.02)$, %, where K_r – instrument reading.
6.3.6	Programmable high- frequency signal generator Г4-164	

Table 2

1	2	3
6.3.7	Programmable high-frequency signal generator Г4-164	
6.3.8	Accepted power wattmeter M3-56 Power amplifier УМ100-400	Power measurement range 0÷20 W in the range up to 17.85 GHz. $\delta_{rel} = \pm\{6 + 0.1[(P_k/P_x) - 1]\}$, %.
6.3.9	Programmable high-frequency signal generator Г4-164 Low-frequency signal generator Г3-112/1 cymometer Ч3-54	
6.3.10	Low-frequency signal generator Г3-112/1 Universal digital voltmeter B7-34	
6.3.11	Distortion meter C6-12 Low-frequency signal generator Г3-112/1 (2 pieces)	Distortion factor measurement range: 0.03÷100 %. $\Delta_{abs} = \pm (0.05 K_r + 0.02)$, %, where K_r – instrument reading, %
6.3.12	Voltmeter checking instrument B1-13	Power measurement limits from 100 μ V to 100 B. $\Delta_{abs} = \pm (5 \cdot 10^{-5} \cdot U_f + 500)$, μ V.

2.2 The use of other check instruments and methods not mentioned in this document may also be admitted if they are approved for use in Russian Federation in accordance with established procedure. Their accuracy rating and characteristics must be no worse than those of the above-mentioned ones.

2.3 All measuring instruments mentioned in the table 2 must be checked in accordance with established procedure and have valid check certificates.

3 SAFETY REQUIREMENTS

3.1 Tester checking is allowed only by persons who have passed the safety briefing.

3.2 Before switching on the tester, the compliance of output voltage from power supply unit with tester operating voltage must be checked.

3.3 Safety requirements stated in the in-line documentation for measuring instruments must be satisfied during the check.

3.4 DO NOT do cable overcommutation while instruments are switched on.

4 CHECK CONDITIONS

4.1 The following conditions should be observed during the check:

- ambient temperature, °C 20 ± 5 ;
- relative air humidity, % $50 \div 80$;
- atmospheric pressure, kPa (mm hg) $86.6 \div 106.7 (650 \div 795)$;
- supply voltage, V 220 ± 22 ;
- supply voltage frequency, Hz 50 ± 1 .

Supply line voltage should be stable and step-free.

5 CHECK PREPARATION

5.1 The following cleanup activities should be carried out before checking the testers:

1) checkable testers and check instruments must be kept indoors under the conditions specified in 4.1 for 2 hours.

2) check instruments must be prepared in accordance with in-line documentation.

3) checkable testers must be prepared in accordance with operations manual.

6 CHECKING

6.1 External examination

6.1.1 Make sure the completeness of the tester PCT-430 in accordance with the tester certificate.

6.1.2 Make sure the conformity of identification mark according to the in-line documentation.

6.1.3 Make sure that tester has no visible damages and contaminations, which can affect its operational capability.

6.1.4 Examine operation and tuning controls for operability (buttons, toggle switches, etc.).

6.1.5 Check for sharp QCD and verification officer stamp impression or state verification certificate (while periodic check).

6.1.6 Checking is not allowed if damages of the tester are revealed.

6.2 Testing

6.2.1 Testing must be carried out in order to check tester's operability. The following operations should be carried out for testing:

1) Make sure that power supply button is in switch-off position.

2) Connect tester to EPU БПЦ15-2.

3) Connect tester to testable radio station as shown in figure 1.




Figure 1

4) Press the power-on button. The following static picture will appear on the screen.




Figure 2

5) Tester will be ready to work in 10 seconds (self-test time).

6.3 Metrological examination

6.3.1 Output signal frequency forming relative error estimation (high-frequency).

6.3.1.1 Use spectrograph C4-74 for signal frequency measuring (high-frequency) in the range from 90 to 210 MHz.

6.3.1.2 Connect tester ("RFI" connector) to spectrograph C4-74 ("Input" connector) using interface cable.

6.3.1.3 Adjust the following parameters of the spectrograph:

- swath = 5 kHz;
- bandwidth = 300 Hz;

Adjust the following parameters of the tester:

- "Receiver" mode
- maximum level (high-frequency)
- turn off modulation
- frequency $F_{hf} = 150$ MHz..

6.3.1.4 Measure signal frequency (high-frequency) using spectrograph C4-74 in accordance with its in-line documentation.

6.3.1.5 Relative error of signal frequency forming (high-frequency) is calculated as:

$$\delta_{rel} = [(F_{yct} - F_{изм}) / F_{yct}] \cdot 100, \%$$

where F_{yct} is setpoint value of output signal frequency (high-frequency) on the tester, $F_{изм}$ is measured value of output signal frequency (high-frequency).

6.3.1.6 Repeat the measurements as described in sections 6.3.1.2 ÷ 6.3.1.5 for frequencies of 90 and 210 MHz.

6.3.1.9 Follow the instructions given in section 6.3.1.5.

The measurement results are satisfactory if the relative error of frequency measurement in the range from 90 to 210 MHz is no more than $\pm 3 \cdot 10^{-4} \%$.

6.3.2 Error of output voltage setting (high-frequency) is estimated following the scheme presented in figure 3.




Figure 3

6.3.2.1 Wire up as shown in Figure 3.

6.3.2.2 Measurement of output voltage level (high-frequency) should be performed using spectrograph C4-74 for alternating voltage with frequency = 90...210 MHz.

6.3.2.3 Adjust the following parameters of the generator Г4-164 in accordance with its in-line documentation:

- voltage level = -110 dB;
- frequency = -150 MHz;
- continuous frequency generation.

6.3.2.4 Energize spectrograph C4-74 by generator Г4-164.

6.3.2.5 Standardize spectrograph C4-74 through input sign with voltage = -120 dB in accordance with its in-line documentation

6.3.2.6 Adjust the following parameters of the tester:

- "Receiver" mode
- voltage level $U_{hf} = -60$ dB;
- deviation = 0 kHz;
- frequency $F_{hf} = 150$ MHz..

6.3.2.7 Measure output voltage (high-frequency) using spectrograph C4-74 in accordance with its in-line documentation.

6.3.2.8 Absolute error of measurement of output voltage signal level (high-frequency) can be defined by the formula:

$$\Delta_{abs} = (U_{hf\ agr} - U_{hf\ m}), \text{ dB},$$

where $U_{hf\ agr}$ is setpoint value of output voltage level (high-frequency) on the tester, dB;

$U_{hf\ m}$ is measured value of output voltage level on the spectrograph C4-74.

6.3.2.9 Repeat the measurements as described in sections 6.3.2.3 ÷ 6.3.2.8, gradually installing frequency = 90 and 210 MHz on the generator Г4-164 and tester.

3.2.10 Repeat the measurements as described in sections 6.3.2.3 ÷ 6.3.2.9, gradually installing voltage level = -80 dB and -60 dB on the generator Г4-164 and tester.

Measurement results are considered to be satisfactory if the absolute error of measured output voltage level (high-frequency) is no more than value defined by the formula: $\Delta_{abs} = \pm [3 + (U_{hf\ agr} / 40)]$, dB.

6.3.3 Relative error of signal frequency deviation setting (high-frequency) is estimated following the scheme presented in figure 4.




Figure 4

6.3.3.1 Wire up as shown in Figure 4.

6.3.3.2 Adjust the frequency of 150 MHz on the microvoltmeter SMV8.5 under its in-line documentation and tune a maximum of declinations

6.3.3.3 Adjust the following parameters of the measuring instrument CK3-45 in accordance with its in-line documentation:

- "FM", "+" mode
- bandwidth $0,2 \div 200$ kHz.

6.3.3.4 Adjust the following parameters of the tester:

- "Receiver" mode
- frequency $F_{\text{вч}} = 150$ MHz
- modulation frequency = 1 kHz
- deviation = 3 kHz.

6.3.3.5 Measure signal frequency deviation (high-frequency) by the measuring instrument CK3-45.

6.3.3.6 Relative error of frequency deviation can be defined by the formula:

$$\delta_{\text{отн}} = \pm [(D_{\text{agr}} - D_{\text{M}}) / D_{\text{agr}}] \cdot 100, \%$$

where D_{agr} is setpoint value of frequency deviation on the tester, kHz;

D_{M} is frequency deviation value measured using the measuring instrument CK3-45, kHz.

6.3.3. Repeat the measurements as described in sections 6.3.3.2 ÷ 6.3.3.6 for frequency deviation = 10 and 20 kHz.

Measurement results are considered to be satisfactory if the relative error of signal frequency deviation (high-frequency) is not more than the value defined by the formula:

$$\delta_{\text{rel}} = \pm [5 + 5 (D_f / D_{\text{agr}})], \%$$

where $D_f = 20$ kHz is upper value of frequency deviation setting range;

D_{agr} is setpoint value of frequency deviation, kHz.

6.3.4 Determination of relative error of output signal frequency forming (low-frequency):

6.3.4.1 Connect tester ("AF OUT" connector) to cymometer Ч3-54 ("Input" connector) using interface cable.

6.3.4.2 Adjust the following parameters of the tester:

- "Transmission" mode;
- output voltage $U_{\text{нч}} = 1$ V;
- frequency $F_{\text{нч}} = 1$ kHz.

6.3.4.3 Measure output signal frequency (low-frequency) using cymometer in accordance with its in-line documentation.

6.3.4.4 Relative measurement error of frequency F_{if} can be calculated as:

$$\delta_{\text{rel}} = \pm [(F_{\text{if agr}} - F_{\text{if M}}) / F_{\text{if agr}}] \cdot 100, \%$$

where $F_{\text{if agr}}$ is setpoint value of output signal frequency (low-frequency) on the tester, kHz;

$F_{\text{if m}}$ is output signal frequency (low-frequency) measured using the cymometer Ч3-54, kHz.

6.3.4.5 Repeat the measurements as described in sections 6.3.4.2 ÷ 6.3.4.4 for tester frequencies = 5 kHz; 20 kHz.

Results are satisfactory if signal frequency error (low-frequency) is not more than ± 1 Hz.

6.3.5 Absolute error estimation of output voltage setting (low-frequency):

6.3.5.1 Connect tester ("AF OUT" connector) and voltmeter B7-34 using interface cable.

6.3.5.2 Adjust the following parameters of the tester:

- "Transmission" mode;
- output voltage $U_{\text{if}} = 0,25\text{V}$;
- frequency $F_{\text{if}} = 1$ kHz.

6.3.5.3 Measure output voltage $U_{\text{нч}}$ using voltmeter B7-34 in accordance with its in-line documentation.

6.3.5.4 Absolute error of output voltage measuring can be defined by the formula:

$$\Delta_{\text{abs}} = (U_{\text{if agr}} - U_{\text{if m}}), \text{ V},$$

where $U_{\text{if agr}}$ is setpoint value of output voltage on tester, V;

$U_{\text{if m}}$ is measured value of output voltage using the voltmeter B7-34, V.

6.3.5.5 Repeat the measurements as described in sections 6.3.5.2 ÷ 6.3.5.4 for output voltage values on the tester = 1 V; 2,0 V.

6.3.5.6 Repeat the measurements as described in sections 6.3.5.2 ÷ 6.3.5.5 for tester frequency $F_{\text{if}} = 5$ kHz; 20 kHz.

Results are satisfactory if absolute error of measured output voltage (low-frequency) does not exceed the value defined by formula:

$$\Delta_{\text{abs}} = \pm (0.02 + 0.05 \cdot U_{\text{if agr}}), \text{ V},$$

where $U_{\text{if agr}}$ is setpoint value of output voltage (low-frequency), V.

6.3.6 **Relative** error of signal frequency (high-frequency) is estimated following the scheme on figure 5.

6.3.6.2 Adjust the following parameters of generator Г4-164 according to its in-line documentation:

- voltage = 1 V;
- frequency $F_{\text{hf}} = 150$ MHz..

6.3.6.3 Adjust "Transmission/ Frequency" mode on the tester.

6.3.6.4 **Take the tester reading.**

6.3.6.5 Relative error of signal frequency F_{hf} (high-frequency) is calculated by the formula:

$$\delta_{\text{rel}} = \pm [(F_{\text{hf agr}} - F_{\text{hf m}}) / F_{\text{hf agr}}] \cdot 100, \%$$

6.3.6.6 Repeat the measurements as described in sections 3.6.2 ÷ 6.3.6.5 for frequency = 90 MHz; 210 MHz on the generator Г4-164.

Measurement results are satisfactory if relative error of the measured signal frequency (high-frequency) does not exceed $\pm 3 \cdot 10^{-4} \%$.

6.3.7 Relative error of signal frequency deviation measuring (high-frequency) is estimated following the scheme in figure 5.

6.3.7.2 Adjust the following parameters of the generator Г4-164 according to its in-line documentation:

- voltage level = 1V
- frequency $F_{\text{Bч}} = 150$ MHz.
- frequency modulation: internal
- deviation = 1 kHz.

6.3.7.3 Adjust "Transmission/ Deviation" mode on the tester.

6.3.7.4 **Take the tester reading.**

6.3.7.5 Relative error of signal frequency deviation (high-frequency) is defined by the formula:

$$\delta_{rel} = \pm [(D_{agr} - D_M) / D_{agr}] \cdot 100, \%$$

6.3.7.6 Repeat the measurements as described in sections 6.3.7.2 ÷ 6.3.7.5 for frequency = 90 MHz; 210 MHz on the generator Г4-164.

6.3.7.7 Repeat the measurements as described in sections 6.3.7.2 ÷ 6.3.7.6 for deviation = 5.0 kHz; 20 kHz on the generator Г4-164.

Measurement results are satisfactory if relative error of measured signal frequency deviation (high-frequency) does not exceed the value defined by the formula:

$$\delta_{rel} = \pm [5 + 5 (D_f / D_M)], \%$$

where $D_f = 20$ kHz is upper value of frequency deviation setting range;

D_M is measured value of frequency deviation, kHz.

6.3.8 **Relative** error of output signal power measuring (high-frequency) is estimated following the scheme in figure 5.




Figure 5

6.3.8.1 **Wire up as shown in Figure 5.**

6.3.8.2 Adjust signal level of the generator Г4-164 satisfying YM100-400 amplifier input power:

- frequency $F_{hf} = 150$ MHz

- power = 20 W.

6.3.8.3 Connect wattmeter M3-56 to power amplifier YM100-400 in accordance with its in-line documentation and measure power amplifier output parameters.

6.3.8.4 Connect tester instead of wattmeter M3-56.

6.3.8.5 **Take the tester reading.**

6.3.8.6 Relative error of power measuring is defined by the formula:

$$\delta_{rel} = \pm [(P_{agr} - P_M) / P_{agr}] \cdot 100, \%$$

6.3.8.7 Repeat the measurements as described in sections 6.3.8.2 ÷ 6.3.8.6 for power values 0.2 W and 10 W.

6.3.8.8 Repeat the measurements as described in sections 6.3.8.2 ÷ 6.3.8.7 for frequency = 90 MHz and 210 MHz..

Measurement results are satisfactory if relative error of power measuring does not exceed the value defined by the formula:

$$\delta_{rel} = \pm [5 + 0.1 \cdot (P_f / P_M)], \%$$

where $P_f = 20$ W is upper value of output signal power measuring range (high-frequency);

P_M is setpoint value of output signal power (high-frequency), W.

6.3.9 Absolute error of signal frequency (low-frequency) is estimated following the scheme in the figure 6.




Figure 6

6.3.9.1 Wire up as shown in Figure 6.

6.3.9.2 Adjust the following parameters of the generator Г3-112/1 in accordance with its in-line documentation:

- output voltage $U_{if} = 1 \text{ V}$;
- frequency $F_{if} = 1 \text{ kHz}$.

6.3.9.3 Check generator frequency using cymometer Ч3-54 in accordance with its in-line documentation.

6.3.9.4 Adjust "Frequency, Vetka" mode on the tester.

6.3.9.5 Take the tester reading.

6.3.9.6 Absolute error of signal frequency F_{if} (low-frequency) is estimated by the formula:

$$\Delta_{abs} = F_{if \text{ agr}} - F_{if \text{ M}}, \text{ kHz},$$

where $F_{if \text{ agr}}$ is setpoint value of frequency, kHz;

$F_{if \text{ M}}$ is measured value of frequency, kHz.

6.3.9.7 Repeat the measurements as described in sections 6.3.9.2 ÷ 6.3.9.6 for frequency = 10 kHz and 100 kHz on the generator Г3-112/1.

Measurement results are satisfactory if absolute error of signal frequency measuring (low-frequency) does not exceed the value defined by the formula:

$$\Delta_{abs} = \pm 1 \text{ Hz}.$$

6.3.10 Relative error of alternating voltage measuring is estimated following the scheme shown in figure 7.




Figure 7

6.3.10.1 Wire up as shown in Figure 7.

6.3.10.2 Adjust "Receiver/ Voltage" mode on the tester.

6.3.10.3 Adjust the following parameters of the generator Г3-112/1:

- frequency $F_{if} = 20 \text{ Hz}$;
- output voltage $U_{if} = 0.2 \text{ V}$.

6.3.10.4 Check generator output voltage by voltmeter B7-34 in accordance with its in-line documentation.

6.3.10.5 Take the tester reading.

6.3.10.6 Relative measurement error of alternating voltage U_{if} is defined by the formula:

$$\delta_{rel} = \pm [(U_{if\ agr} - U_{if\ m}) / U_{if\ agr}] \cdot 100, \%$$

where $U_{if\ agr}$ is setpoint value of generator alternating voltage, V,
 $U_{if\ m}$ is alternating voltage value measured by the tester, V.

6.3.10.7 Repeat the measurements as described in sections 6.3.10.2 ÷ 6.3.10.6 for generator frequencies = 1.0 kHz and 20 kHz.

6.3.10.8 Repeat the measurements as described in sections 6.3.10.2 ÷ 6.3.10.7 for generator output voltage = 1.0 V and 15 V.

Measurement results are satisfactory if relative error estimation of input voltage (low-frequency) does not exceed the value **defined by the formula:**

$$\delta_{rel} = \pm [3 + 3 (U_f / U_m)], \%$$

where $U_f = 20$ V is **upper value of alternating voltage measuring range,**

U_m is measured value of alternating voltage, V.

6.3.11 **Relative error of alternating voltage** distortion factor is estimated following the scheme shown in figure 8.




Figure 8

6.3.11.1 **Wire up as shown in Figure 8.**

6.3.11.2 Adjust "Receiver/ **distortion factor**" mode on the tester.

6.3.11.3 Adjust the following parameters of the generator Г3-112/1 (Г1):

- frequency $F_{if} = 1$ kHz;

- output voltage $U_{if} = 1$ V.

6.3.11.4 Adjust the value of distortion factor of **5%** on the instrument C6-12 changing output voltage of the generator Г3-112/1 (Г2).

6.3.11.5 **Take the tester reading.**

6.3.11.6 **Relative error estimation of distortion factor** is defined by the formula:

$$\delta_{rel} = \pm [(DF_{agr} - DF_m) / DF_{agr}] \cdot 100, \%$$

where DF_{agr} is setpoint value of distortion factor on the instrument C6-12, %,
 DF_m is distortion factor measured by the tester, %.

6.3.11.7 Repeat the measurements as described in sections 6.3.11.2 ÷ 6.3.11.6 for distortion factor of 20 % and 50 % on the instrument C6-12 .

6.3.11.8 Repeat the measurements as described in sections 6.3.11.2 ÷ 6.3.11.6 for Г3-112/1 (Г1) generator frequency of 3 kHz.

Measurement results are satisfactory if relative error of distortion factor **in the range from 1 to 50 %** does not exceed the value defined by the formula:

$$\delta_{rel} = \pm [5 + 0.1 \cdot (DF_f / DF_m)], \%$$

where $DF_f = 50$ % is **upper value of distortion factor measuring range, %;**

DF_M is measured value of distortion factor, %.

Note: Measurement error of distortion factor in the range from 50 to 100 % is not standardized.

6.3.11 Relative measurement error of direct voltage is estimated as described below:

6.3.11.1 Connect the tester ("AF IN" connector) to voltmeter B1-12 ("Output" connector) using interface cable.

6.3.11.3 Adjust voltage of 0.1 V on the voltmeter B1-13 in accordance with its in-line documentation.

6.3.11.3 Adjust "Receiver" mode on the tester.

6.3.11.4 Take the tester reading.

6.3.11.5 Relative error estimation of direct voltage is defined by the formula:

$$\delta_{OTH} = \pm [(U_{agr} - U_M) / U_{agr}] \cdot 100, \%$$

where U_{agr} is voltmeter value of direct voltage, V;

U_M is direct voltage value measured by the tester, V.

6.3.11.6 Repeat the measurements as described in sections 6.3.11.2 ÷ 6.3.11.5 for direct voltage of 0.3 V, 3 B and 20 V on the voltmeter.

Measurement results are satisfied if direct voltage does not exceed the value defined by the formula:

$$\delta_{rel} = \pm [2 + 0.05 \cdot (U_f / U_M)], \%$$

where $U_f = 30$ V is upper value of direct voltage measuring range,

$U_{изм}$ is measured value of direct voltage, V.

7 CHECK RESULT PRESENTATION

7.1 Check results of testers should be **protoled on form of mandatory** appendix B.

7.2 Check certificate is **given out** if check results are satisfactory.

7.3 Issue and use of testers is not allowed if they do not pass primary check. In this case **trouble report is given out**.

7.4 Testers must be removed from use **and repaired if** they do not pass periodical check.

APPENDIX A
(for reference)

Metrological performance of tester PCT-430

Name of characteristic, unit	Parameter	Error limits
1 Output signal frequency range (high-frequency), MHz	90...210	$\delta_{rel} = \pm 3 \cdot 10^{-4} \%$ for the range
2 Output level range (high-frequency), dB. (0 dB = 1 mW)	From - 60 to - 130	$\Delta_{abs} = \pm [3 + (U_{hf\ agr} / 40)]$, dB.
3 Signal frequency deviation range (high-frequency), kHz.	0.2 ÷ 20	$\delta_{rel} = \pm [5 + 5 (D_f / D_{agr})]$, %, where $D_f = 20$ kHz is upper value of deviation frequency setting range; D_{agr} is setpoint value of deviation frequency, kHz.
4 Output signal range (low-frequency), kHz.	0.02 ÷ 20	± 1 Hz
5 Output voltage range (low-frequency), V.	0.02 ÷ 2	$\Delta_{a\delta c} = \pm (0.02 + 0.05 \cdot U_{lf\ agr})$, V, where $U_{lf\ agr}$ is setpoint value of output voltage, V.
6 Output voltage distortion factor (low-frequency), %.	No more than 1	—
7 Output signal frequency range (high-frequency), MHz.	90...210	$\delta_{rel} = \pm 3 \cdot 10^{-4} \%$.
8 signal frequency deviation measuring range (high-frequency), kHz.	0.2 ÷ 20	$\delta_{rel} = \pm [5 + 5 (D_f / D_{agr})]$, %, where $D_f = 20$ kHz is upper value of deviation frequency setting range; D_{agr} is measured value of deviation frequency, kHz.
9 Output signal power measuring range (high-frequency), W.	0.2 ÷ 20	$\delta_{rel} = \pm [5 + 0,1 \cdot (P_f / P_M)]$, %, where $P_f = 20$ W is upper value of output signal power measuring

10 Signal frequency measuring range (low-frequency), Hz.	$20 \div 1 \cdot 10^5$
11 Sinusoidal alternating voltage measuring range in the frequency range from 0.02 to 20 kHz, V.	$0.02 \div 15$

range (high-frequency); P_M is **measured value** of output signal power (high-frequency), W.

$$\pm 1 \text{ Hz.}$$

$$\delta_{\text{rel}} = \pm [3 + 2 (U_f / U_M)], \%$$

where $U_f = 20 \text{ V}$ is **upper** value of alternating voltage measuring **range**; U_M is **measured value** of alternating voltage, V.

APPENDIX A

Name of characteristic, unit	Parameter	Error limits
12 Measuring range of alternating voltage distortion factor, %.	1 ÷ 100	<p>In the range from 1 to 50 %:</p> $\delta_{rel} = \pm [5 + 0,1 \cdot (DF_f / DF_M)], \%$ <p>where $DF_f = 50 \%$ is upper value of distortion factor measuring range, %; DF_M is DF measured value, %.</p> <p>In the range from 50 to 100 % DF is not standardized.</p>
13 Direct voltage measuring range, V.	0.02 ÷ 20	$\delta_{rel} = \pm [2 + 2 (U_f / U_M)], \%$ <p>where $U_f = 30 \text{ V}$ is upper value of direct voltage measuring range; U_M is direct voltage measured value, V.</p>

APPENDIX B
(mandatory)

PCT-430 TEST LOG

(works number)

owned by _____
(company name)

checked by _____
(company name)

“ _____ ”

Check conditions

Ambient temperature, °C _____

Relative humidity, % _____

Atmospheric pressure, kPa (mm hg) _____

Supply voltage, V _____

Check instruments

1 External examination, completeness checking _____

Conclusion: _____

2 Testing _____

Conclusion: _____

3 Measurement of tester parameters, errors

Table A.1

Item number	Measurable parameter	Error limits

Conclusion: _____

Overall conclusion: _____

(certificate, number or cause of unserviceability)

Chief of State Supervision Laboratory _____
(signature)

(surname)

State verification officer _____
(signature)

(surname)

Stamp here

Date _____

